



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/616,102	07/09/2003	Hans Schmotzer	MEISS71.007DV1	1655
20995	7590	01/05/2006	EXAMINER	
KNOBBE MARTENS OLSON & BEAR LLP			PELLEGRINO, BRIAN E	
2040 MAIN STREET			ART UNIT	
FOURTEENTH FLOOR			PAPER NUMBER	
IRVINE, CA 92614			3738	

DATE MAILED: 01/05/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

6

**Advisory Action  
Before the Filing of an Appeal Brief**

Application No.

10/616,102

Applicant(s)

SCHMOTZER ET AL.

Examiner

Brian E. Pellegrino

Art Unit

3738

**--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**

THE REPLY FILED 22 December 2005 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE.

1. ☒ The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:

- a) ☒ The period for reply expires 3 months from the mailing date of the final rejection.  
b) ☐ The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.

Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**NOTICE OF APPEAL**

2. ☐ The Notice of Appeal was filed on \_\_\_\_\_. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).

**AMENDMENTS**

3. ☒ The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because  
(a) ☒ They raise new issues that would require further consideration and/or search (see NOTE below);  
(b) ☐ They raise the issue of new matter (see NOTE below);  
(c) ☐ They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or  
(d) ☒ They present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: See Continuation Sheet. (See 37 CFR 1.116 and 41.33(a)).

4. ☐ The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).  
5. ☒ Applicant's reply has overcome the following rejection(s): 112 2<sup>nd</sup> rejection.  
6. ☐ Newly proposed or amended claim(s) \_\_\_\_\_ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).  
7. ☒ For purposes of appeal, the proposed amendment(s): a) ☒ will not be entered, or b) ☐ will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.  
The status of the claim(s) is (or will be) as follows:  
Claim(s) allowed: \_\_\_\_\_  
Claim(s) objected to: \_\_\_\_\_  
Claim(s) rejected: 1-2/  
Claim(s) withdrawn from consideration: \_\_\_\_\_

**AFFIDAVIT OR OTHER EVIDENCE**

8. ☐ The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e).  
9. ☐ The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing of good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).  
10. ☐ The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached.

**REQUEST FOR RECONSIDERATION/OTHER**

11. ☒ The request for reconsideration has been considered but does NOT place the application in condition for allowance because:  
See Continuation Sheet.  
12. ☐ Note the attached Information Disclosure Statement(s). (PTO/SB/08 or PTO-1449) Paper No(s). \_\_\_\_\_  
13. ☒ Other: the objection to the specification has been overcome.

Continuation of 3. NOTE: the new limitation in claim 1 that the "pre-determined distance is defined by a template having a distance between a bore and a contact surface of the template.... requires further consideration.

Continuation of 11. does NOT place the application in condition for allowance because: Applicant argues that D'Antonio does not disclose the "pre-determined distance from a femoral condyle". However, it is the Examiner's position that the prior art sizing device has a "pre-determined distance defined by the template used to prepare the femur. Additionally it should also be noted that claims 3,9,20,23 all claim the pre-determined distance is "about 5%- 15%". Since the specification fails to define what the deviation is by using the term "about 5%" it can be interpreted that this deviation can be  $\pm 5\%$ . Since the term "about" is used in the claims, it can be said that "about 5%" can mean about 0 or a minimal difference between the amount of bone resected and the distance between the condyle engaging surface and the peg of the femoral implant. Clearly D'Antonio discloses at a minimum a template that has a "pre-determined" distance and when the femur is resected can be done with a little more (col. 1, lines 53-64) such that it is "about 5% more" or "about 0" than a dimension of the slideway implant.



**BRIAN E. PELLEGRINO**  
PRIMARY EXAMINER